



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: June 15, 2005 INITIALS: DEC OWNER: WSP
DISTRICT: 3 ADDRESS: Highway 24 MP 43
LOCATION: Vernita #48 COUNTY: Benton
MAKE: Rix Lake TYPE: Truck scale ACCURACY CLASS: TL
MODEL: LQ 310 ID. NO: 972046 CAPACITY: 100,000 lbs SECTIONS: 2
MINIMUM GRADUATIONS: 20 lb REG. ELEMENT: electronic PLATFORM: 8' x 10'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Center</u>	<u>4000</u>	<u>4000</u>	<u>0</u>				
	<u>8000</u>	<u>8000</u>	<u>0</u>	<u>Sect 1</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>
	<u>12000</u>	<u>12000</u>	<u>0</u>				
	<u>16000</u>	<u>16000</u>	<u>0</u>	<u>Sect 2</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>
<u>Center</u>	<u>20000</u>	<u>20000</u>	<u>0</u>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 22205 - 7 and 8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by:

David Cronin

Date:

June 16, 2005

State of Washington

County of Thurston

Signed and sworn to (or affirmed) before

me on June 16, 2005 by

Michael J. Evans

(Signature of notary public)

Michael J. EVANS

Printed Name of Notary

My appointment expires 12-27-07

